



[Redacted]	Replace	--	
Issue date	2022 September, 28	Scope	EMC / Medical
[Redacted]	Appendix	--	

TEST VERIFICATION OF CONFORMITY

Kiwa hereby declares that the PTZ Camera, types:

CM7*-IP-*

CM70-IP-W (20 x optical zoom, white version),

CM71-IP-W (12 x optical zoom, white version),

CM73-IP-W (30 x optical zoom, white version),

CM70-IP-B (20 x optical zoom, black version),

CM71-IP-B (12 x optical zoom, black version),

CM73-IP-B (30 x optical zoom, black version)

manufactured by: Avonic B.V.

have been tested to following standards:

EN 60601-1-2 (2015), class A

EN 60601-1-2 (2015), home healthcare environment and
professional healthcare facility environment

and fulfills the requirements as mentioned in the test report.

D. van der Vlugt
Business Unit Manager
Kiwa Dare